

AzuraQuant[™] Green 1-Step qPCR Mix LoRox

Catalog No.	Pack Size and Concentration	Components and Volume	
AZ-3701	100 x 20 µl reactions, 2x	2x 1-Step qPCR Mix LoRox - 1 x 1 mL	20x RTase 1 x 0.1 mL
AZ-3705	500 x 20 µl reactions, 2x	2x 1-Step qPCR Mix LoRox - 5 x 1 mL	20x RTase 5 x 0.1 mL
AZ-3710	1000 x 20 µl reactions, 2x	2x 1-Step qPCR Mix LoRox - 10 x 1 mL	20x RTase 10 x 0.1 mL
AZ-3750	5000 x 20 µl reactions, 2x	2x 1-Step qPCR Mix LoRox - 10 x 5 mL	20x RTase 10 x 0.5 mL

Description

The AzuraQuant[™] Green 1-Step qPCR Mix is a ready-to-use 2x master mix and companion thermostable reverse transcriptase for use in highly sensitive SYBR[™] real-time RT-PCR assays. The AzuraQuant[™] Green 1-Step qPCR Mix is powered by Azura[™] HS Taq DNA Polymerase, AzuraSprint[™] Reverse Transcriptase, and an optimized buffer chemistry including a non-inhibitory intercalating dye for robust first-strand cDNA synthesis and real-time PCR in a single tube. The Mix delivers early quantification cycle values (Ct) and broad range detection for increased sensitivity, speed and reproducibility. The AzuraQuant[™] Green 1-Step qPCR Mix can be used to quantify a specific target RNA from either total RNA or mRNA while reducing the number of pipetting steps and time to result. The 2x mix also contains an optimized dUTP concentration which provides for the optional use of DNA glycosylase to control carry-over contamination. In order to determine instrument compatibility and the most appropriate formulation, please refer to the qPCR Reagent Compatibility Tool.

Storage

AzuraQuantTM Green 1-Step qPCR Mix LoRox is shipped on blue ice and should be stored at -20° C upon receipt. Excessive freeze/thawing should be avoided. When stored as specified, AzuraQuantTM Green 1-Step qPCR Mix LoRox is stable for 12 months from date of receipt.

Important Guidelines

- Use primer-design software, such as Primer3 (http://frodo.wi.mit.edu/primer3/) or visual OMP[™] (http://dnasoftware.com/). Primers should have a melting temperature (Tm) of approximately 60°C.
- Optimal amplicon length should be 80 bp-200 bp, and should not exceed 400 bp.
- Different real-time PCR instruments require different levels of ROX[™] passive reference dye. Generally, modern instruments do not require passive reference but include the option to use it for normalization. Please refer to the qPCR Reagent Compatibility Tool to determine which kit is the most suitable for your instrument.
- When comparing AzuraQuant Green 1-Step qPCR Mix with a reagent from an alternative supplier, we strongly recommend
 amplifying from a 10-fold template dilution series. Loss of detection at low template concentration is the only direct
 measurement of sensitivity. An early Ct value is not an indication of good sensitivity, but rather an indication of reaction
 speed.

Reaction setup

1. Prepare a qPCR master mix based on following table. Briefly mix AzuraQuant[™] Green 1-Step qPCR Mix by inversion:

Component	20µl Reaction	Final Cor	centration/Notes
AzuraQuant [™] Green 1- Step qPCR Mix LoRox	10 µl	1x	
Forward Primer (10 μ M)	0.8 µl	400 nM	
Reverse Primer (10 μ M)	0.8 µl	400 nM	
20x RTase	1.0 µl	1x	Add prior to RNA
Template RNA	10 pg to 100 ng Total RNA >0.01pg mRNA	variable	
PCR-grade water	Up to 20 μl final volume		

* For alternative total reaction volumes (eg. 25 µl), scale all components proportionally to maintain final concentrations.

2. Program the qPCR instrument using following conditions, acquiring data on the appropriate channel:

Cycles	Temperature & Time	Notes
1	50°C, 10 minutes	cDNA synthesis
1	95°C, 2 minutes	Enzyme activation
30 - 40	95°C, 5 seconds 60°C to 65°C, 20 - 30 seconds	Denaturation Anneal/Extension (do not exceed 30 seconds and do not use temps below 60°C

Quality Control

AzuraQuant[™] Green 1-Step qPCR Mix LoRox is tested extensively for robust activity, processivity, efficiency, heat activation, sensitivity, absence of nuclease contamination and absence of nucleic acid contamination. AzuraQuant[™] Green 1-Step qPCR Mix LoRox is manufactured under a comprehensive quality management system, following ISO 9001:2015 standards.

Limitations of Use

This product is intended for research purposes only and is not intended for any animal or human therapeutic use.

Technical Support

For Trouble-shooting and Technical Guidance, please contact us at <u>tech@azuragenomics.com</u> and provide qPCR reaction conditions, cycling parameters, amplicon size, and screen grabs (amplification traces and melting profiles) if possible.

Azura Genomics guarantees the performance of all products in the manner described in our product literature. The purchaser must determine the suitability of the product for its particular use. Azura Genomics shall not be liable for any direct, indirect, consequential or incidental damages arising from the use, the result of use or the liability to use this product.